

ThermaLite Food Probe Thermometer







Labfacility are the Tiles leading manufacturer of Temperature Sensors, Thermocouple Connectors and associated Temperature Instrumentation and stockings of Thermocouple Cables. The Company has been trading since 1971 and is ISO9001 accredited.





ThermaLite instant-read thermometer for day-to-day food and catering applications.

Fast and accurate readings

High accuracy with a wide temperature range

CalCheck 0.0 °C Function

Free traceable certificate of calibration

Waterproof to

IP66

Polygiene BioMaster

product protection

Meets the European Standard EN 13485

Ideal for HACCP compliance, the ThermaLite is a simple and accurate pocket thermometer.

The ThermaLite has a long battery life of 2,000 hours. It will switch off automatically after 10 minutes to maximise battery life and indicate when the battery is running low. If required, this food thermometer probe has the option to toggle between °C/°F with ease.

The reduced-tip stainless steel food probe ensures accurate readings. And with a diameter of just 2 mm, you'll hardly see where the measuring tip has been inserted.

The thermometer is supplied with a protective probe cover to ensure it is kept safe during storage.

Specifications

Specifications

| Product Code | XE-4203-001 |
|------------------|--|
| Accuracy | +/-0.5 °C (-10 to 100 °C), +/-2.5 °C (200 to 300 °C) otherwise +/-1.5 °C |
| Resolution | 0.1°C / F |
| Battery | 3 volt CR2032 lithium coin cell |
| Colour | White |
| Dimensions | 22 x 36.3 x 223 mm |
| Display Type | 25 x 23.5 mm LCD |
| Max. Temperature | 300°C |



| Min. Temperature | -40°C |
|-----------------------|---|
| Weight | 67 Grams including probe cover |
| Power Consumption | 2000 hour battery life |
| Sensor Type | Thermistor |
| Standards Met | IP66 |
| Accessories | FREE traceable calibration certificate |
| Construction | ABS plastic includes Biomaster product protection |
| Guarantee | 1 year |
| Important Information | Please note that certificates supplied do not display an expiry date. They state the date the item was manufactured. The frequency of recalibration is determined by each customer's specific application requirements and internal quality procedures. |